

Certificate FI20/966583

Cherry medical holding (shenzhen) Co.,Ltd.


Room 602, Yuanling Building, No.2002, Hongli Road,
Nantian Community,
Yuanling Street,
Futian District,
Shenzhen,
China

It is certified that the manufacturer's technical file and the PPE product detailed on
page 2 have been assessed and found to be in accordance with

Regulation (EU) 2016/425 Module B, EU type-examination

This certificate is valid from 15 September 2020 until 15 September 2025
1. Certified since 15 September 2020

Authorised by





FINAS
Finnish Accreditation Service
S003 (EN ISO/IEC 17065)

SGS FIMKO OY, Notified Body 0598

Takomatie 8, FI-00380, Helsinki, Finland
t +358 9 696 361 f +358 9 692 5474 www.sgs.com

Page 1 of 2



This document is issued by the Company subject to its General Conditions of
Certification Services accessible at www.sgs.com/terms_and_conditions.htm.
Attention is drawn to the limitations of liability, indemnification and jurisdictional
issues established therein. The authenticity of this document may be verified at
<http://www.sgs.com/en/certified-clients-and-products/certified-client-directory>.
Any unauthorized alteration, forgery or falsification of the content or appearance
of this document is unlawful and offenders may be prosecuted to the fullest
extent of the law.

Cherry medical holding (shenzhen) Co.,Ltd.

Regulation (EU) 2016/425

Module B, EU type-examination

Issue 1

PPE Product

CHERRY (logo) QRYW202005 particle filtering half mask, consisting of a white five layer (polypropylene/poly tetra fluoroethylene) disposable face mask, with aluminium nose clip, and polyester/spandex ear loop with plastic hook.

It is certified that the manufacturer's technical file and the above mentioned PPE have been assessed and found to meet the applicable Essential Health and Safety Requirements in Annex II of Regulation (EU) 2016/425 Personal Protective Equipment

The following have been applied:

EN 149:2001+A1:2009 (Respiratory protective devices - filtering half masks to protect against particles) device classification: FFP2 NR.

This certificate is issued on the strict condition that appropriate checks on manufactured PPE, as detailed in Article 19 (c) of the Regulation are implemented and maintained while the model is in production

Certification is based on technical file reference:

Filtering half mask/ QRYW202005, version 1, dated: 2020-08-04.

SGS Reference Number UK/CRS 241718.

This certificate remains the property of SGS Fimko Oy to whom it must be returned on request